AUTOMATIC ELMENDORF - DIGITAL TEAR TESTER

ELMATIC





COMPANY WITH MANAGEMENT SYSTEM CERTIFIED BY DNV GL = ISO 9001 = = ISO 14001 =

ELMATIC code 275D

Automatic Elmendorf - digital tear tester

Main features

- Automatic digital tear tester, suitable for testing specimens within a very wide tear range, like textiles, paper, leather, plastic, non-woven and other materials
- Measuring range up to 30.000 cN
- · Automatic test execution
- · Automatic pendulum restoring to the starting position
- · Total safety protective cover with auto-locking system
- LCD display
- PC & PRINTER connection
- · High accuracy and repeatability of results
- High efficiency



Automatism

Test execution

After sample insertion (1), press START to perform testing; the test sequence (from 2 to 7) will be automatically executed in total safety, according to the following steps:

- 1. manual insertion of sample
- 2. pre-cut of specimen
- 3. pendulum release
- 4. specimen tear
- 5. measurement of tear force
- 6. pendulum braking
- 7. pendulum raising to its starting position

This automatism ensures high testing efficiency, high accuracy, and repeatability of results.

Pendulum restoring

- · Automatic detection of pendulum's weight
- · Automatic restoring of the starting position



1. Manual insertion of sample




4. Specimen tear




6. Pendulum braking



2. Specimen pre-cut



5. Measurement of tear force



7. Pendulum raising to its starting position

Safety

ELMATIC is designed in accordance with CE safety regulation.

Safety cover with auto-locking system

Thanks to its safety cover, the access of the operator to the testing area is possible only if the instrument is in safe conditions; in fact, both specimen cut and pendulum swing are enabled only when the safety cover is perfectly closed; the auto-locking system allows its opening only if the instrument is not working.

Pendulum front loading

By pressing the "F1" key, pendulum automatically moves to the front loading position, so that the operator can open the cover and easily place the weights in the correct position. As soon as the cover is closed, by pressing the "START" key, the pendulum is released and then automatically raised to the starting position again. Warning message in case of incorrect pendulum positioning.

Setting and results

Keyboard and LCD display are designed for use even in the production area.

The user-friendly design enables intuitive selection of functions and easy setting of the test parameters, such as:

- selection of the testing standard from a pre-loaded list, or input of different test parameters and specific tolerance range;
- input of test references, such as: company name, measuring unit, selected weight, number of sample's layers, direction (weft / warp, transversal / longitudinal).

The printout will show additional information, such as the "out-of-range" results.

Connections

Connection to printer

To print test results and statistical data, a mini thermal printer is available as optional accessory.

Connection to PC

ELMATIC software (already included) enables the data transfer to PC for further graphic and statistical elaboration, as well as to print and save test reports.

		*********	******	TEST PARAME	TERS					ST	ATISTIC	CAL RESUL	LTS				MESD
*MESDANLAB			Company name: Sample code:		MESDANLAB oY209			MI	Min N			256.43			OTHER DRAFT AND IN		
Measure Unit : cN Pendulum (cN) : 3200									Max N			487.49			and the second sec		
				Measure unit:		cN 3200 cN			Me	Mean CV[%] Sigma			391.1150 15.7076			07/03/2010 SAVE	
				Pendulum:					CV								
Standard EN 1S013937-1			International st			EN ISO13937-1			Sig				61.4348		RECALL		
			Test direction: WEFT/TRAMA			4			Range [%]				59.0773			PRINT	
		15 -85 %		and the second se													TXT
irec	tion test :	WEFT															
				TEST	TEARI	NG	487.49-										
EARING FORCE STATISTIC			3	416.	42	460.00-											
			2	256.		440.00-											
	61.43	Ran	59.08	3	397.		420.00-										
				4	364. 446.		400.00-		-								
ean	391.12	Cv2	15.71		371.		380.00-				-						
in	256.43	Max	487.49	7	371.		360.00-										
					423.		340.00-										
- REJECT TEST RESULT				375.		320.00-											
				10	487.	49	300.00-										
it o	f range stan	dard		12			280.00-										
				13			260.00-										
	TEARING FORC	E		3.4			\$ 240.00-										
	416.42	-		15			₩ 220.00-										
-				16			200.00-										
2	256.43			10			160.00-										
3	397.75			19			140.00-										
4	364.18			20			120.00-										
5	446.33			21			100.00-										
2				22			80.00-										
0	371.64			23			60.00-										
7	371.64			25			40.00-										
8	423.90			26			20.00-										
0	375.37			27			0.00-										
				28			0	1 2 3	4 5	6 7 6	9 10 11				21 22 23 2	24 25 26 2	7 28 29 30 3
10	487.49			29			Tear		Mean	_	-	Numi	ber of test				
		Elmatic	MESDAN LAB	30		_											



ELMATIC

TECHNICAL DATA

- · measuring pendulum range: 1.600 3.200 6.400 15.000 30.000 cN
- · unit of measurement: cN N daN mN gr kg lbs oz
- · clamps opening: Max 4,5 mm
- \cdot cut length: 20 mm ± 1
- · tear length: 43 mm
- \cdot tolerance range: 20-80% of each full scale, settable range or full scale
- \cdot angular resolution: digital optical encoder with resolution 0.09 degrees
- · combinable pendulum weights: 2 weights
- · number of samples clamped together at the same time for testing (only for paper)
- · display: 128x64 LCD display
- \cdot blade: tungsten steel
- \cdot data port: 2 RS232 ports for PC and printer connection
- · language: Italian-English
- · statistics: force min. max. mean, CV%, St. Deviation, Range%

DIMENSIONS / POWER SUPPLY

Weight: 66 kg Dimensions: (L) 510 x (W) 700 x (H) 630 mm Power supply: 115 up to 230 Vac, 50/60 Hz, single-phase

STANDARD ACCESSORIES

WEIGHTS KIT complete with 2 combinable weights in order to obtain as measuring range 1.600-3.200-6.400-15.000- 30.000cN		275D.300
SPECIMEN PREPARATION KIT complete with one template for cutting, and one cork base	code	275D.310
one spare blade	code	275D.52
ELMATIC CD software		

MATIC CD software

OPTIONAL ACCESSORIES

CALIBRATION KIT complete with 4 check weights	code	275D.400
set of 5 spare blades	code	275D.402
mini thermal printer	code	275D.136
calibration report	code	275D.CC1

REFERENCE STANDARDS

TEXTILE - according to: UNI EN ISO 13937-1, UNI EN ISO 4674-2 (coated fabrics), ASTM D 1424, ASTM D 751 (coated fabrics), UNI EN ISO 1974, M&S P29 PLASTIC - according to: UNI EN ISO 6383-2-method 360A PAPER - according to: TAPPI T414, UNI EN ISO 1974

Photographs and descriptions of the present leaflet have to be considered as purely indicative and not binding

